


CERTIFICATO DI TARATURA LAT 146 13175
Certificate of Calibration

- data di emissione <i>date of issue</i>	2021/05/26
- cliente <i>customer</i>	SPECTRA S.r.l. Via J. F. Kennedy, 19 - 20871 Vimercate (MB)
- destinatario <i>receiver</i>	Alter Eco s.a.s. Via F.lli Cuzio, 42 - 27100 Pavia (PV)
- richiesta <i>application</i>	T319/21
- in data <i>date</i>	2021/05/17
<u>Si riferisce a</u> <i>referring to</i>	
- oggetto <i>item</i>	Fonometro
- costruttore <i>manufacturer</i>	LARSON DAVIS
- modello <i>model</i>	831
- matricola <i>serial number</i>	0002151
- data di ricevimento oggetto <i>date of receipt of item</i>	2021/05/20
- data delle misure <i>date of measurements</i>	2021/05/26
- registro di laboratorio <i>laboratory reference</i>	21-0745-RLA

Il presente certificato di taratura è emesso in base all'accreditamento LAT N° 146 rilasciato in accordo ai decreti attuativi della legge n. 273/1991 che ha istituito il Sistema Nazionale di Taratura (SNT). ACCREDIA attesta le capacità di misura e di taratura, le competenze metrologiche del Centro e la riferibilità delle tarature eseguite ai campioni nazionali e internazionali delle



essere riprodotto salvo espressa autorizzazione del Centro.

This certificate is issued in accordance with the accreditation LAT N° 146 which is connected with the International System of Units (SI) and has established the traceability of the measurements to the International System of Units (SI).

ACCREDIA attests the measurement and calibration capabilities and the metrological competence of the Centre and the traceability of the measurements to the International System of Units (SI).

Partially reproduced, without the permission of the Centre.

Refer to page 2 of this certificate and the respective reference standards which are valid in the laboratory.

Refer to page 2 of this certificate and the related calibration certificate.

I risultati di misura riportati nel presente Certificato sono stati ottenuti in base ai campioni di prima linea dei certificati di taratura, in corso di validità. Essi si riferiscono esclusivamente al momento e nelle condizioni di taratura, salvo diversamente specificato. The measurement results reported in this Certificate were obtained following the use of primary reference standards are indicated as well, from which starts the traceability chain. They relate only to the calibrated item and they are valid for the time and conditions of calibration, unless otherwise specified.

Le incertezze di misura dichiarate in questo documento sono state determinate conformemente al documento EA-4/02 e sono espresse come incertezza estesa ottenuta moltiplicando l'incertezza tipo per il fattore di copertura k corrispondente ad un livello di fiducia di circa il 95%. Normalmente tale fattore k vale 2.

The measurement uncertainties stated in this document have been determined according to EA-4/02. They were estimated as expanded uncertainty obtained multiplying the standard uncertainty by the coverage factor k corresponding to a confidence level of about 95%. Normally, this factor k is 2.

 Il Responsabile del Centro
 Head of the Centre

